

A NEW NONINVASIVE METHOD OF CARDIAC OUTPUT DETERMINATION. PRINCIPLE OF MEASUREMENT

Artur Szczepański, Jacek Jastrzębski, Tadeusz Pałko, Andrzej Książkiewicz, Maciej Guć

Abstrac

We present the principles of an indicator method in which the indicator is either hypertonic salt solution or distilled water injected as a bolus into the venous system. The dilution curve is measured in temporal or limb artery by means of external reography.

Keywords: cardiac output, indicator dilution techniques, electric impedance